

Band Life and Reuses

To ensure the integrity of the band, do not exceed the following guidelines:

- Length of band usage - 2 weeks under typical conditions
- Number of reuses - up to 8

Note: Aggressive attempts by the patient to remove the tag may reduce band life. Re-use of the band more than 8 times may render it unreliable. If you have any doubt, dispose of the old band and apply a new one.

Reference Documentation

For more information about any of these procedures, see the "Pedz User Guide" (p/n 0980-317-000).

Battery Information

The Pedz Tag uses one non-replaceable battery (Model CR2032 or equivalent, electrical rating 3V, 220mAh).

The Pedz system software automatically displays a Low Battery notification if the Pedz Tag enters a low battery state during use.

Handling Information

Battery Handling - this device contains a lithium battery. Do not force open, heat to 212°F (100°C), or dispose of in fire.

X-rays - Do not directly expose the tag to X-rays. (The tag is not affected by stray radiation.)

MRI - this tag contains metal components and is not approved for use near MRI.

Warranty

STANLEY Healthcare's products are warranted against defects in materials and workmanship and shall perform in accordance with published specifications for a period of one year. STANLEY Healthcare Solutions reserves the right to change specifications without notice.

Limitation of Liability

This Product has been designed for use to assist in the prevention of pediatric abductions.

The range, accuracy, function and performance of this Product may vary from the published specifications due to many factors, including, but not limited to, site impairments from structural effects, metal objects in the vicinity, placement of the receiver and transmitter, interference from other electrical devices, atmospheric effects, installation, and maintenance. There may be other factors, which also affect performance of this Product.

STANLEY Healthcare ("STANLEY") does not guarantee that this Product will detect 100% of pediatric abductions, nor does STANLEY guarantee that this Product will be free from returning false reports of pediatric abductions.

Monthly testing and maintenance of this Product, as described in the Product documentation, is essential to verify the system is operating correctly and to ensure that the probability of detecting an alarm and/or locating the transmitter are maximized.

The failure to undertake regular testing and maintenance will increase the risk of system failure and failure to detect pediatric abductions. The failure to undertake regular testing and maintenance will increase the risk of false reports of pediatric abductions.

STANLEY hereby disclaims all warranties, express or implied, arising out of or in connection with any of its Products of the use or performance thereof, including but not limited to, where allowable by law, all other implied warranties or conditions of merchantability, fitness for a particular purpose, and those arising by statute or otherwise in law or from a course of dealing or usage of trade.

STANLEY's liability to you or anyone claiming through or on behalf of you with respect to any claim or loss arising out of the use or misuse of STANLEY's Product, defective products or materials, improper

installation or maintenance of STANLEY's Product or products or the system in which they are incorporated, or alleged to have resulted from an act or omission of STANLEY or any person, negligent or otherwise, shall be limited to the following, at STANLEY's sole option:

- 1) **the repair or replacement of defective Product or materials supplied by STANLEY during the warranty period as set out in the Product documentation; or, at the option of STANLEY,**
- 2) **a refund of the purchase price of the Product supplied by STANLEY.**

In no event shall STANLEY be liable for general, specific, indirect, consequential, incidental, exemplary or punitive damages or any losses or expenses suffered by you or anyone else, whether or not STANLEY, or its employees, officers, agents, resellers or installers has been informed of the risk of such loss or expense and whether or not such losses or expenses were foreseeable.

Statements

United States—Federal Communication Commission (FCC)

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. NOTE: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense. **Warning: Changes or modifications not expressly approved by STANLEY could void the user's authority to operate the equipment.**

Canada—Industry Canada

The term "IC:" before the certification/registration number only signifies that Industry Canada technical specifications were met.

Conforms to
UL STD 294



Telephone: +1-800-824-2996
Web Site: www.stanleyhealthcare.com
© 2017 STANLEY Healthcare. All rights reserved. April 2017, 0980-326-000 Rev A, K#H: 10096
Part Number: 8066A7001

Pedz
Pediatric Protection

Pedz Tag
User Sheet

Instructions for the Pedz™ Tag

Applying the Tag

The tag is not sterilized at the factory and should be cleaned before first use. The Pedz Tag may be applied to either the wrist or ankle as per hospital policy.

- 1 Measure the child's extremity using the Sizing Guide.
- 2 Use the Band Clippers to cut the Banding Material to obtain a cut-to-length band.
- 3 Firmly push one end of the band into either of the openings on the side of the tag until you feel the locking mechanism engage.



- 4 Wrap the band around the extremity.
- 5 Firmly push the loose end of the band into the opposite opening on the side of the tag until you feel the locking mechanism engage. A flashing light on the front of the tag indicates a successful connection.



- 6 Optionally, for children over three years of age, snap a Tag Skin onto the face of the tag.



- 7 Verify that the tag has been admitted into the system at the Pedz computer.

Assessing Tag Fit

When you apply a tag to the child, ensure that a secure fit is obtained. A secure fit is defined as a fit snug enough to prevent the tag from being slipped off the hand/heel, and loose enough to wiggle but not rotate freely.

- 8 It is recommended that a staff member examine each Pedz Tag and tag site regularly (per hospital policy) for appropriate fit and the absence of any skin irritation.
- 9 If the Pedz Tag appears to be too tight, as evidenced by tag / band impressions on the child's skin or the inability to move the tag on the skin surface, remove the tag immediately and apply a new band.

Removal Key

The Removal Key is used by authorized staff to quickly release the band from the tag. It can be attached to a lanyard.



Removing the Tag

- 1 If there is a Tag Skin, remove it by lifting one edge and pulling it away from the tag's face.
- 2 Slide the Removal Key over the face of the tag from either side (for left or right-handed use).



- 3 Apply pressure using your thumb and forefinger to release the end of the band from the tag.



- 4 Discard the band if the child is being discharged. Return the tag for cleaning and reuse.

Cleaning the Tag and Removal Key

Consult with your Infection Control representative for cleaners available to your hospital that have been approved for use on plastics as stated in STANLEY Healthcare Technical Bulletin [KB# 8184] "Approved Cleaning Materials for Patient Security Tags" – available on the STANLEY Healthcare Support Community (stanleyhealthcare.com/support).

- 1 Use a germicidal wipe to remove heavy soil.
- 2 Spray the cleaner onto the surface and openings.



- 3 If necessary, a soiled tag or Removal Key can also be soaked in a cleaning solution.



- 4 Clean the surface and openings with a small brush, as needed.



- 5 Rinse with water and dry with a clean cloth or allow to air dry.